

02/12/2014



Pressure Mapping Test Results in **Harmony**

Tilt & Space Recliner Chair


with **Reflection** Foam

(Aged: Male, mid 40's, weight approx. 13 stone/82 Kilos)



WITHOUT Treat Eezi Pad (2 layer)



Minimum (mmHg)	0.00
Maximum (mmHg)	53.70
Average (mmHg)	27.04
Variance (mmHg ²)	96.28
Standard deviation (mmHg)	9.81
Coefficient of variation (%)	36.29
Horizontal center (cm)	22.05
Vertical center (cm)	22.71
Sensing area (cm ²)	1431.38
Regional distribution (%)	100.00



WITH Treat-Eezi Pad (2 layer).



Minimum (mmHg)	0.00
Maximum (mmHg)	37.47
Average (mmHg)	22.40
Variance (mmHg ²)	61.21
Standard deviation (mmHg)	7.82
Coefficient of variation (%)	34.92
Horizontal center (cm)	20.55
Vertical center (cm)	22.43
Sensing area (cm ²)	1601.21
Regional distribution (%)	100.00

