

Care and Maintenance

- Wash periodically by hand with a damp cloth and mild soap
- Use a towel to absorb most of the dampness and then allow to dry at room temperature.
- Do not hang up, or iron and do not expose to direct heat sources such as stoves, heaters, radiators, direct sunlight etc
- Do not expose to alcohol, ointments or solvents.



High Arm Sling

Fitting Guidelines & Care Instructions



Version 2
October 2013

CHANECO
4 Kilvey Road, Brackmills, Northampton, NN4 7BQ
T: 01604 709999 F: 01604 700057
www.chaneco.co.uk

Chaneco trading as Chris Hanley & Partners which is registered in England as a limited liability partnership under number LP7980



Single Patient Use

Not
Fire
Retardant



Indications

Indications for use:


- Post surgery
- Hand surgery
- Post trauma

Fitting Instructions


The High Arm Sling is universal for both left and right arms. These instructions for use demonstrate the product in a right arm application, however the left arm configuration is an identical application process.



1. Place the injured arm into the sling, ensuring the elbow is held securely in place.




2. Whilst holding the sling in a closed position, attach the securing strap to the front of the sling as shown and fold over to secure.



3. To attach the D-ring strap, ensure it is positioned 5-7cms up from the patients wrist. Fold over to secure, ensuring that the side facing D-ring is positioned on the inside of the sling.




4. Place the shoulder strap over the uninjured shoulder and feed through the D-ring attachment, securing the strap firmly with the Velcro® fastening as shown. Adjust the height as required.




PLEASE NOTE:
To use as a shoulder immobiliser - please follow these additional instructions:

1. Attach the immobilisation strap to the front of the High Arm Sling, 3-5cms away from the securing strap.



2. Place the immobilisation strap around the back of the waist.



3. Feed the immobilisation strap through the side facing the D-ring (as shown) and secure with the Velcro® fastening